

BIOCOMPATIBILITY CERTIFICATE

Testmaterial: Feinpolyamid PA 2200
Chemical composition: Polylaurinlactam

Supplier: EOS GmbH; Electro Optical Systems
Robert-Stirling-Ring 1, 82152 Krailling; Germany

Studies performed: The following study was performed in order to determine the biocompatibility of the device.

USP BIOLOGICAL TESTS (CLASSIFICATION VI/121 °C)

Results: The test item did not show any effect in the performed study and meets all the criterias of the USP Biological Tests Classification VI.

BSL BIOSERVICE Scientific Laboratories GmbH

Behringstraße 6

D-82152 Planegg


Dr. Ingrid Haist
Biological Safety Testing
Date: December 13, 2004

